

Listing#: E115723  
Report #: 126829B  
Original Certification Date: August 9, 2023  
Revised Certification Date:

This Certification is issued to:  
Concens A/S  
Oddesundvej 1, 6715 Esbjerg N  
Denmark

Stating that the product(s):  
Actuator Controller,  
Models C3 – CON-0W, C4 – CON-xW

Product Rating(s):  
•Model C3 – CON-0W: 24V  
•Model C4 – CON-xW: 24V

Achieved Certification to the following standard(s):  
AAMI ES60601-1:2005 (R2012) + A1:2012 / CSA C22.2 No. 60601-1:14 (R2018), 3rd Ed. Medical  
Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance



Hon Keung Ip  
Eurofins Electrical and Electronic Testing North America, Inc.

*All changes proposed in the previously identified product that affects the above information must be submitted to Eurofins for evaluation prior to implementation to assure continued NRTL Certification status. The covered product(s) shall be subject to follow-up inspections to ensure that the Certified product(s) are identical to the product sample evaluated by Eurofins E&E NA and that all responsibilities are being fulfilled as specified in the Applicants' Responsibility section of the Certification Report. The Applicant named above has been authorized Eurofins E&E NA to represent the product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the Eurofins E&E NA Applicant Contract, Listing Reports, and the applicable agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in this Certification program. This certification has been granted under a System 3 program as defined in ISO/IEC 17067.*



*Eurofins E&E North America, Inc. is accredited by OSHA and the Standards Council of Canada.*

**NRTL**